



Fluid Precision since 1991

aperainst.com

2310-C High-Range Conductivity Electrode

User Manual



Brief Introduction

Apera Instruments 2310-C High-Range Conductivity Electrode is featured with a proprietary platinum black ring sensor for accurate readings in an ultra-wide conductivity range up to 2000 mS/cm.

Technical Specifications

Range	20 to 2000 mS/cm
Housing	POM
Constant	K=10
Connector	BNC
Cable length	3 Ft
Dimension	ø12*145 mm
Temperature sensor	N/A
Operating temperature	32 to 176°F (0 - 80°C)

How to Install the Electrode

1. Find the conductivity BNC socket (where it shows Cond) on the conductivity meter; Open the rubber cap; Insert the BNC connector of the electrode to the BNC socket while twisting clockwise until it's locked.
2. After connecting the new electrode to your meter, adjust the meter's constant K to 10 according to the instruction of your meter's manual. Then perform a 2 to 3-point calibration to ensure the accuracy.

How to Use the Electrode

1. There is a proper amount of deionized water in the cap on top of the electrode. The conductivity sensor is soaked in it to keep its sensitivity.
2. Before measuring, loosen the bottle cap, then pull out the electrode while twisting counterclockwise. Place the storage bottle at a safe position.

3. Rinse the electrode with distilled or deionized water and blot-dry it with clean tissue or filter paper. Rinse the electrode again with your test solution if possible for the best result. Never rub the sensor.
4. Insert the electrode into your sample solution and stir it for a few seconds in the solution to remove potential air bubbles. Then wait for the stable reading and take the measurement.
5. After measurement, insert the electrode back into the storage cap and twist it on. If the water in the cap is contaminated, rinse out the storage cap and fill in new deionized or distilled water.

How to Maintain the Electrode

1. Always rinse the electrode with distilled or deionized water before and after each test and calibration.
2. Store the electrode in the storage cap filled with distilled or deionized water when not in use.
3. Keep the electrode connector clean and dry. Use cotton balls with isopropyl alcohol to clean if it gets dirty and then blow-dry it. This is to prevent a potential short circuit, which will undermine the electrode's performance.

Limited Warranty

We warrant this electrode to be free from defects in material and workmanship and agree to repair or replace free of charge, at the option of APERA INSTRUMENTS, LLC, any malfunctioned or damaged product attributable to responsibility of APERA INSTRUMENTS, LLC for a period of SIX MONTHS from the delivery.

This warranty does not cover any damages due to:

Accidental damage, transportation, storage, improper use, failure to follow the product instructions, unauthorized repair, normal wear and tear, or any other actions or events beyond our reasonable control.

APERA INSTRUMENTS, LLC

Address: 6656 Busch Blvd, Columbus, OH 43229

Phone: +1 (614) 285-3080

Website: aperainst.com

Email: info@aperainst.com